



TECHNOZYM® vWF:CBA ELISA

Calibrators and Controls

LOT RB47C00

3015987-180

Table • Tabelle • tabella • tabla • tabela • tabell

DIL 0.5mL AQUA

REF CALIBRATORS	LOT		Sollwert/ Nominal value	Kalibrierung durch®, nach Calibration by®, acc, to	
vWF CAL 1 TH10001ND	1ND31C0	2025-08-31	1.34 IU/mL	IS 07/316	
vWF CAL 2 TH10002ND	2ND21C1	2026-04-30	1.02 IU/mL	IS 07/316	
vWF CAL 3 TH10003ND	3ND21C1	2026-04-30	0.51 IU/mL	IS 07/316	
vWF CAL 4 TH10004ND	4ND22C1	2026-04-30	0.28 IU/mL	IS 07/316	
vWF CAL 5 TH10005ND	5ND22C1	2026-04-30	0.00 IU/mL	IS 07/316	

REF CONTROLS	LOT		Sollwert/ Nominal value	Vertrauensbereich Confidence range	Kalibrierung durch®, nach Calibration by®, acc, to
vWF CONT H TH1000AND	AND31C1	2026-04-30	1.05 IU/mL	0.79 – 1.31 IU/mL	IS 07/316
vWF CONT L TH1000BND	BND21C1	2026-04-30	0.42 IU/mL	0.25 – 0.59 IU/mL	IS 07/316

