


Table • Tabelle • tabella • tabla • tabela • tabell

DIL 1mL AQUA

TECHNOVIEW Arixtra® Control High **LOT** 3V02C00

REF	LOT		Sollwert Nominal value	Vertrauensbereich Confidence range	Kalibrierung* Calibration
TECHNOVIEW Arixtra CONT H 5090032	3V02C00	2022-11-30	1.41 µg/mL	1.06 – 1.76 µg/mL	INTERNAL REFERENCE STANDARD

* mit Technochrom® anti-Xa am Ceveron® alpha
using Technochrom® anti-Xa on Ceveron® alpha

Qualitätskontrolle
quality control

