


# Technoview LMWH CON H

**LOT** 3H22C00

 **2024-08-31**

3015002- 024 02/11/2022



Test test	Reagenz Reagent	Einheit unit	Richtwert (Vertrauensbereich) expected value (confidence range)	Std.
Technochrom Anti Xa	CON H	IU/mL	1.17 ( 0.88 - 1.46 )	IS 11/176