


Table • Tabelle • tabella • tabla • tabela • tabell

DIL 1mL AQUA

TECHNOVIEW® Rivaroxaban\* Calibrator Set High LOT 0381B00

REF	LOT		Sollwert Nominal value	Kalibrierung** Calibration
TECHNOVIEW Rivaroxaban CAL 1 High 5090180	5181B00	2020-06-30	0.0 ng/mL	HPLC-MS/MS
TECHNOVIEW Rivaroxaban CAL 2 High 5090182	5281B00	2020-06-30	47.9 ng/mL	HPLC-MS/MS
TECHNOVIEW Rivaroxaban CAL 3 High 5090184	5381B00	2020-06-30	103.3 ng/mL	HPLC-MS/MS
TECHNOVIEW Rivaroxaban CAL 4 High 5090186	5481B00	2020-06-30	219.6 ng/mL	HPLC-MS/MS
TECHNOVIEW Rivaroxaban CAL 5 High 5090188	5581B00	2020-06-30	466.4 ng/mL	HPLC-MS/MS

\* Rivaroxaban ist die aktive Substanz aus Xarelto®, welches eine registrierte Marke der Firma Bayer ist  
Rivaroxaban is the active substance in Xarelto® which is a registered trade mark of Bayer

\*\* mit Technochrom® anti-Xa am Ceveron® alpha  
using Technochrom® anti-Xa on Ceveron® alpha

Qualitätskontrolle  
quality control