



Table • Tabelle • tabella • tabla • tabela • tabell

DIL 0.5mL **AQUA**

REF CALIBRATORS	LOT		Sollwert/ Nominal value	Kalibrierung durch [®] , nach Calibration by [®] , acc, to
vWF CAL 1 TH10001ND	1ND32C1	2026-10-31	1.68 IU/mL	IS 07/316
vWF CAL 2 TH10002ND	2ND31C1	2026-10-31	0.83 IU/mL	IS 07/316
vWF CAL 3 TH10003ND	3ND31C1	2026-10-31	0.39 IU/mL	IS 07/316
vWF CAL 4 TH10004ND	4ND31C1	2026-10-31	0.19 IU/mL	IS 07/316
vWF CAL 5 TH10005ND	5ND31C1	2026-10-31	0.00 IU/mL	IS 07/316

REF CONTROLS	LOT		Sollwert/ Nominal value	Vertrauensbereich Confidence range	Kalibrierung durch [®] , nach Calibration by [®] , acc, to
vWF CONT H TH1000AND	AND32C1	2026-10-31	0.96 IU/mL	0.72 – 1.20 IU/mL	IS 07/316
vWF CONT L TH1000BND	BND31C1	2026-10-31	0.41 IU/mL	0.25 – 0.57 IU/mL	IS 07/316

