


Table • Tabelle • tabella • tabla • tabela • tabell

DIL 1mL AQUA

TECHNOVIEW® Rivaroxaban\* Control Low **LOT** 3411C00

<b>REF</b>	<b>LOT</b>		Sollwert Nominal value Technochrom® anti-Xa	Vertrauensbereich Confidence range Technochrom® anti-Xa	Sollwert Nominal value Coamatic Heparin	Vertrauensbereich Confidence range Coamatic Heparin	Kalibrierung durch®, nach Calibration by®, acc, to
TECHNOVIEW Rivaroxaban <b>CONT L</b> 5090190	3411C00	2023-04-30	68.3 ng/mL	47.8 – 88.8 ng/mL	83.1 ng/mL	58.2 – 108.0 ng/mL	INTERNAL REFERENCE STANDARD

\* Rivaroxaban ist die aktive Substanz aus Xarelto® welches eine registrierte Marke der Firma Bayer ist  
Rivaroxaban is the active substance in Xarelto® which is a registered trade mark of Bayer

Qualitätskontrolle  
quality control



1/1

