

Ceveron TGA RC High Kit

LOT **B513C00**

 **2022-09-30**

3013832- 009 08/11/2021

Ceveron 100 series barcode



Kontrolle control	Lot	Parameter	Einheit unit	Richtwert (Vertrauensbereich) expected value (confidence range)	Std.
Ceveron TGA Control High	1Y13C00	Peak thrombin	nM	479 (359 - 599)	IS 01/580
Ceveron TGA Control Low	2712C00			185 (139 - 231)	IS 01/580