


**DIL** 1.0 mL Aqua dest.

**TECHNOZYM<sup>®</sup> Vitronectin Standard**

<b>LOT</b>		<b>Nominal value Sollwert µg/mL</b>	<b>calibrated against / kalibriert gegen</b>
<b>RDR61B4</b>	<b>2025-05-31</b>	<b>405.3</b>	<b>Internal reference plasma</b>

**Standard Dilution:**

<b>Tube</b>		<b>Incubation buffer</b>	<b>Conc. %</b>	<b>Conc. µg/mL</b>
Predilution	0.1 mL Standard	0.9 mL	10	40.53
A	0.01 mL Predilution	0.99 mL	0.1	0.405
B	0.5 mL from A	0.5 mL	0.05	0.203
C	0.5 mL from B	0.5 mL	0.025	0.101
D	0.5 mL from C	0.5 mL	0.0125	0.051
E	-	0.5 mL	0	0

Qualitätskontrolle  
quality control

