



# Declaration of Conformity



Doc. Nr.: DoC9G.01

p. 1 | 1

Temp.: TP0484\_04  
Att.: 01 to 000100

<b>Manufacturer</b>	Technoclone Herstellung von Diagnostika und Arzneimitteln GmbH Brunner Str. 67 1230 Vienna Austria		
<b>SRN</b>	<b>AT-MF-000024115</b>		
<b>Product(s)</b>	Citrate Buffer		
<b>Basic UDI-DI</b>	912003686TC9G000000144		
<b>UDI-DI (GTIN)</b>	<b>Product name</b>		<b>REF</b>
09120036860858	Citrate Buffer	60 mL	5400045
09120036860865	Citrate Buffer	25 mL	5400047
<b>Classification</b>	Class A according to Annex VIII rule 5(a)		
<b>Common Standards</b>	N/A		
<b>Notified Body Involvement</b>	N/A		
<b>Conformity assessment route</b>	Following article 48(10), the EU declaration of conformity is issued after drawing up the technical documentation set out in Annexes II and III (IVDR)		
<b>Initial date of first declaration of conformity</b>	01-Feb-2023		
<b>Valid until</b>	31-Jän-2026		

This declaration of conformity is issued under the sole responsibility of Technoclone Herstellung von Diagnostika und Arzneimitteln GmbH. We hereby declare that the device(s) specified above meet the provision of the Regulation for *in vitro* diagnostic medical devices (EU) 2017/746 Annex I.

**Name:** Veronika Binder

**Position:** Chief Executive Officer

**Signed:**

**Date:** 01-Feb-2023

This declaration of conformity is issued in Vienna, Austria on behalf of Technoclone Herstellung von Diagnostika und Arzneimitteln GmbH.