

Mit 0.5 mL Aqua dest. rekonstituieren / reconst. with 0.5 mL Aqua dest.

Artikel / Article CALIBRATORS	Lot / batch	Ablaufdatum / expiry date	Sollwert / Nominal value Adamts - 13 Activity	Sollwert / Nominal value Adamts - 13 Antigen
Adamts - 13 Calibrator 1 Art.Nr.: 5450610	1RT81B1	2021-01-31	0.79 IU/mL	0.88 IU/mL
Adamts - 13 Calibrator 2 Art.Nr.: 5450612	2RT81B1	2021-01-31	0.41 IU/mL	0.48 IU/mL
Adamts - 13 Calibrator 3 Art.Nr.: 5450614	3RT81B1	2021-01-31	0.20 IU/mL	0.23 IU/mL
Adamts - 13 Calibrator 4 Art.Nr.: 5450616	4RT81B1	2021-01-31	0.10 IU/mL	0.12 IU/mL
Adamts - 13 Calibrator 5 Art.Nr.: 5450618	5RT81B1	2021-01-31	0.00 IU/mL	0.00 IU/mL

Artikel / Article CONTROLS	Lot / batch	Ablaufdatum / expiry date	Sollwert / Nominal value		Vertrauensbereich Confidence range
Adamts - 13 High Control Art.Nr.: 5450620	RRY81B1	2021-01-31	Antigen	0.71 IU/mL	0.53 – 0.89 IU/mL
			Activity	0.59 IU/mL	0.44 – 0.74 IU/mL
Adamts - 13 Low Control Art.Nr.: 5450622	RRZ81B1	2021-01-31	Antigen	0.18 IU/mL	0.11 – 0.25 IU/mL
			Activity	0.16 IU/mL	0.10 – 0.22 IU/mL

IU/mL wurden gegen den WHO 1<sup>st</sup> international standard ADAMTS-13 plasma (12/252) kalibriert.  
IU/mL have been calibrated against the WHO 1<sup>st</sup> international standard ADAMTS-13 plasma (12/252)