

TECHNOZYM ADAMTS13 ELISA

LOT RQ42C00

 2026-02-28

3050183- 182 12/08/2024

| Test test | Reagenz reagent | Einheit unit | Sollwert (Activity) Nominal value (Activity) | Sollwert (Antigen) Nominal value (Antigen) | Std. |
|--------------------------|-----------------|--------------|--|--|--------|
| TECHNOZYM ADAMTS13 ELISA | CAL 1 | IU/mL | 1.25 | 1.23 | 12/252 |
| | CAL 2 | IU/mL | 0.58 | 0.53 | |
| | CAL 3 | IU/mL | 0.42 | 0.37 | |
| | CAL 4 | IU/mL | 0.13 | 0.11 | |
| | CAL 5 | IU/mL | 0.00 | 0.00 | |

| Test test | Reagenz reagent | Einheit unit | Sollwert (Activity) Nominal value (Activity) | Sollwert (Antigen) Nominal value (Antigen) | Std. |
|--------------------------|-----------------|--------------|--|--|--------|
| TECHNOZYM ADAMTS13 ELISA | Cont H | IU/mL | 0.70 (0.53 - 0.88) | 0.68 (0.51 - 0.85) | 12/252 |
| | Cont L | IU/mL | 0.13 (0.08 - 0.18) | 0.12 (0.07 - 0.17) | |