

Table • Tabelle • tabella • tabla • tabela • tabell

DIL 0.5mL **AQUA**

REF CALIBRATORS	LOT		Sollwert/ Nominal value	Kalibrierung nach Calibration acc, to
ADAMTS-13 Activity CAL 1 5450791	1UT71B0	2019-06-30	1.14 IU/mL	IS 12/252
ADAMTS-13 Activity CAL 2 5450792	2UT71B0	2019-06-30	0.55 IU/mL	IS 12/252
ADAMTS-13 Activity CAL 3 5450793	3UT71B0	2019-06-30	0.28 IU/mL	IS 12/252
ADAMTS-13 Activity CAL 4 5450794	4UT71B0	2019-06-30	0.13 IU/mL	IS 12/252
ADAMTS-13 Activity CAL 5 5450798	5UT71B0	2019-06-30	0.07 IU/mL	IS 12/252
ADAMTS-13 Activity CAL 6 5450795	6UT71B0	2019-06-30	0.00 IU/mL	IS 12/252

REF CONTROLS	LOT		Sollwert/ Nominal value	Vertrauensbereich Confidence range	Kalibrierung nach Calibration acc, to
ADAMTS-13 Activity CONT H 5450796	RU71B0	2019-06-30	0.78 IU/mL	0.59 – 0.98 IU/mL	IS 12/252
ADAMTS-13 Activity CONT L 5450797	RUZ71B0	2019-06-30	0.22 IU/mL	0.13 – 0.31 IU/mL	IS 12/252

1 IU entspricht 100% ADAMTS-13 Aktivität
1 IU equates to 100% ADAMTS-13 activity